

SIRMAXO CHEMICALS PVT.LTD.

Mr. Gopalakrishna
E – 35, MID C Tarapur

Mumbai, Palghar Dist, Thane
- Indien -

Your sign

Our trials in
KW 31

Our signs
GiS / WHC

Date
26.08.2010

Trials with the manual disinfectant Sanidex OPA

Dear Mr. Gopalakrishna

The test series on your manual liquid disinfectant, a high level disinfectant **Sanidex OPA**, for use in plunge bath, is accomplished.
The tested materials are representative for the materials we use in our products.

Test conditions:

The parts have been deposited 100 cycles each with 40 minutes affecting time in a ready to use solution.

Results:

There have not evolved any damages on KARL STORZ instruments by the use of the disinfectant. This passes for the complete tested KARL STORZ product spectrum.

Hausadresse:
KARL STORZ GmbH & Co. KG
Mittelstraße 8
D-78532 Tuttlingen

Tel.: (07461) 7080
Fax: (07461) 708105
E-Mail: info@karlstorz.de
Web: www.karlstorz.de

Kommanditgesellschaft:
KARL STORZ GmbH & Co. KG
Sitz der Gesellschaft: Tuttlingen
Handelsregister: Tuttlingen HRA 442
Steuer-Nr.: 21062/03455

Persönlich haftender Gesellschafter:
KARL STORZ-Verwaltungsgesellschaft mbH
Sitz der Gesellschaft: Tuttlingen
Handelsregister: Tuttlingen HRB 225
Geschäftsführer: Sybill Storz

Bankverbindungen:
Volksbank Donau-Neckar eG, Tuttlingen (BLZ 643 901 30)
Kto.-Nr. 772 003, SWIFT: GENO DE3303
IBAN: DE97 6439 0130 0000 7720 03

Dresdner Bank AG Tuttlingen (BLZ 643 800 11)
Kto.-Nr. 2 713 305 00, SWIFT: DRES DE 3303
IBAN: DE69 6438 0011 0271 3305 00

Kreissparkasse Tuttlingen (BLZ 643 500 70)
Kto.-Nr. 1322 SWIFT: SOLA DES 1 TUT
IBAN: DE79 6435 0070 0000 0013 22

Deutsche Bank AG Tuttlingen (BLZ 653 700 75)
Kto.-Nr. 211 639 000 SWIFT: DEUT DE3303
IBAN: DE09 6537 0075 0211 6390 00


- 2 -

You can use the name „**KARL STORZ**“ in safety data sheets only under conditions as follows:

1. The composition of your product is unchanged according to ours.
2. Our release is strictly referenced to the product **Sanidox OPA**
3. **The SIRMAXO CHEMICALS PVT LTD.** Is exclusively named as producer of the product
4. You' re pointing out that the compatibility of our instruments is only given in case of following the specified utilization instructions.

We hope we could help you with these informations. For further cooperation do not hesitate to contact us.

With best regards
KARL STORZ GmbH & Co. KG


Gisela Schnellen
Reprocessing Specialist
- Quality Assurance Dept. -